



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED

By Carol Day at 11:40 am, Jul 14, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204118	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 07/12/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Troop G		TIME OF INSPECTION 9:31 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 07/12/2014 21:31	
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR	
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS	
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50°C	<input checked="" type="checkbox"/> QUARTZ STANDARD	
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION	
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER	
<input checked="" type="checkbox"/> INDICATOR LIGHTS		
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc LOT # 13290 EXP. DATE 10/29/2015		
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.1 °C SIMULATOR SN G11115 EXP. DATE 01/09/2015		
<input checked="" type="checkbox"/> CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) <input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE <input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE		
TEST 1 0.095	TEST 2 0.096	TEST 3 0.097

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE Daniel W. Hinten	PRINT FULL NAME Daniel W. Hinten
TYPE II PERMIT NUMBER/EXPIRATION DATE 240240 09/28/2014	TELEPHONE NUMBER (417) 469-3121

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

800 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is October 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



DANIEL W HINTEN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/28/2012

Number 220312

Expires 09/28/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204118
07/12/14
21:31

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 50c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefg h i j k l m n o
p q r s t u v w x y z { | } ~ "

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket


MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204118
07/12/14

ARREST TIME: 11:11
SUBJECT NAME:
SELF
DOB: 11/11/11 SEX: M
STATE/D.L.: MO/
ARRESTING OFFICER:
HINTEN/D/W
OFFICER I.D.: 778
TESTING OFFICER:
SAME
OFFICER I.D.:
PERMIT NUMBER: 240240
EXPIRATION DATE: 05/16/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 21:40
INTERNAL STANDARD VERIFIED 21:40
RADIO INTERFERENCE

Operator Signature



Face This Side ~~Down~~ This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204118
07/12/14

TESTING OFFICER:

HINTEN/D/W
OFFICER I.D.: 778
PERMIT NUMBER: 240240
EXPIRATION DATE: 05/16/16
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	21:34
INTERNAL STANDARD	VERIFIED	21:34
EXTERNAL STANDARD	.095	21:34
BLANK TEST	.000	21:35
EXTERNAL STANDARD	.096	21:35
BLANK TEST	.000	21:36
EXTERNAL STANDARD	.097	21:37
BLANK TEST	.000	21:37

N = 3

SIM. = .1

AVG. = .096

Operator Signature

